

## DEVELOPMENT AND VALIDATION OF RP-HPLC METHOD FOR SIMULTANEOUS ESTIMATION OF IVABRADINE AND NEBIVOLOL IN TABLET DOSAGE FORMS

GK. Rajeswari<sup>1\*</sup>, Siva Srikrishna<sup>2</sup>, T. Priyanka<sup>1</sup>, P. Mounika<sup>1</sup>,  
G. Sadhana<sup>1</sup> and B. Roshini<sup>1</sup>

<sup>1</sup>Department of Pharmaceutical analysis, University College of Technology,  
Osmania University, Telangana-500 007, India.

<sup>2</sup>Department of Pharmaceutics, Samskruti College of Pharmacy,  
Kondapur, Telangana, India.

### ABSTRACT

A new, simple and sensitive Reverse phase performance liquid chromatography (RP-HPLC) method has been developed for the separation and quantification of Ivabradine (IVD) and Nebivolol (NBL) in tablet dosage form. The flow rate was maintained at 1.0ml/min and the eluent was monitored at 260nm. The retention time of IVD and NBL were 1.82min and 4.42 min respectively. The method was validated in terms of linearity, precision, accuracy, specificity and robustness. The method was linear and for precision studies; RSD for IVD and NBL were 0.02 and 0.04 respectively. The percentage recoveries for both drugs from their tablets were 100.80 and 99.76% respectively.

**Keywords:** RP-HPLC, Ivabradine (IVD) and Nebivolol (NBL).

### INTRODUCTION

Analytical chemistry is not a separate branch of chemistry, but simply the application of chemical knowledge. Physico-chemical methods are used to study the physical phenomenon that occurs as a result of chemical reactions. Method for analyzing drugs by HPLC demands primary knowledge about the nature of the sample, structure, polarity, volatility, stability and the solubility parameter. Method validation is the process to confirm that the analytical procedure employed for a specific test is suitable for its intended use. Methods need to be validated or revalidated<sup>1</sup>. Such as Z. Precision, Specificity, Accuracy, Linearity, Range, Limit of detection (LOD), Limit of Quantification (LOQ), Ruggedness.<sup>1-4</sup>

### MATERIALS AND METHODS<sup>5-7</sup>

Waters HPLC system connected with UV- Visible – SPD 10A Vpseries Detector and Empower-2 Software was used. Rheodyne 7725i injection with 20 $\mu$ L loop and analytical column- WATERS XTERRA RP8 4.6x150, 5microns are used. Ivabradine and Nebivolol. were generously given by LARA Drugs Pvt Ltd, Hyderabad, and

Telangana, India. Acetonitrile (HPLC grade) was procured from E.Merck (India) Ltd, Mumbai. Methanol and orthophosphoric acid (AR grade) were procured from Qualigens fine chemicals, Mumbai<sup>2</sup>. Water (HPLC grade) was obtained from a Milli-QRO water purification system.

### Preparation of mobile phase

Accurately measured 400 ml (40%) of above buffer and 600 ml of A triethylamine HPLC (60%) were mixed and degassed in an ultrasonic water bath for 10 minutes and then filtered through 0.45  $\mu$  filter under vacuum filtration.

### Diluent Preparation

The Mobile phase was used as the diluent.

### Preparation of Standard Solution

Accurately weighed amount of 1 mg Ivabradine and 10 mg Nebivolol. Were taken to a 10 ml cleaned and dried volumetric flask. This was then diluted with 7ml of diluent and was sonicated. The volume was made to 10 ml with the same solvent. This was marked and labeled as Stock solution. Further, an amount of 0.3 ml amlodipine and telmisartan each were pipette

from the above stock solution into a 10ml volumetric flask and diluted up to the mark with diluents to get 30 µg/ml of Ivabradine and 15 µg/ml of Nebivolol.

### Preparation of Sample Solution

Accurately weighed amount of 1 mg Ivabradine and 10 mg Nebivolol were taken to a 10 ml cleaned and dried volumetric flask. This was then diluted with 7ml of diluent and was sonicated<sup>3</sup>. The volume was made to 10 ml with the same solvent. This was marked and labeled as Stock solution. Further, an amount of 0.3 ml Ivabradine and Nebivolol each was pipette from the above stock solution into a 10ml volumetric flask and diluted up to the mark with diluents to get 30 µg/ml of Ivabradine and 15 µg/ml of Nebivolol. The standard and sample solution of 30 µg/ml of Ivabradine and 15 µg/ml of Nebivolol was injected for five times and the peak areas were recorded.

### Validation Parameters<sup>8-10</sup>

#### System suitability

The chromatographic systems used for analysis must pass the system suitability limits before sample analysis can commence. Set up the chromatographic system; allow the HPLC system to stabilize for 40 min<sup>4</sup>. Inject blank preparation (single injection) and standard preparation (six replicates) and record the chromatograms to evaluate the system suitability parameters like resolution, tailing factor, theoretical plate count and % RSD for peak area of six replicate injections of LMS standard (% RSD NMT 2.0). The system suitability data is reported in table.

#### Linearity

The linearity of the method was determined at five concentration levels ranging from 10-50 µg/ml for Ivabradine and 5 to 25µg/mL for Nebivolol respectively the linearity was evaluated by linear regression analysis, using least squares method<sup>5</sup>. The slope and intercept value for calibration curve was  $y = 19288X$  ( $r^2 = 0.9996$ ) for Ivabradine and  $y = 16616X$  ( $r^2 = 0.999$ ) for Nebivolol. The results shows that an excellent correlation exists between response

factor and concentration of drugs within the concentration range indicated above. The calibration curves are shown in Table and Figure.

#### Accuracy

For accuracy determination, three different concentrations were prepared separately<sup>6</sup> i.e. 80%, 100% and 120% for the analytic and chromatograms are recorded for the same. The data was given in the table.

#### Intermediate Precision

30 µg/ml of Ivabradine and 15 µg/ml of Nebivolol of the above sample solution were injected for five times in five different days and peak areas were recorded.

Chromatograms were recorded and results are shown in Table.

#### Limit of Detection

The limit of detection (LOD) and limit of quantification (LOQ) of the developed method were determined by injecting progressively low concentrations of the standard solutions using the developed RP-HPLC method<sup>8</sup>. The LOD is the smallest concentration of the analytic that gives a measurable response. The LOD for Ivabradine and Nebivolol found to be 30µg/mL and 15 µg/mL, respectively.

#### Limit of Quantification

The LOQ is the smallest concentration of the analyte, which gives response that can be accurately quantified (signal to noise ratio of 10)<sup>10</sup>. The LOQ was 30µg/mL and 15µg/mL for Ivabradine and Nebivolol, respectively in Table and Figure.

#### Ruggedness

The ruggedness of the method was determined by carrying out the experiment on different instruments like Shimadzu HPLC (LC-2010AHT), Agilent HPLC and Waters' Breeze HPLC by different operators using different columns of similar type like Hypersil C18, Phenomenex Gemini C18 and Hichrom C18. Data is represented in Table.

## RESULTS AND DISCUSSION

### System suitability

**Table 1: Results of system suitability parameters for Ivabradine and Nebivolol**

S.No	Name	Retention time(min)	Area (µV sec)	Height (µV)	USP resolution	USP tailing	USP plate count
1	Ivabradine	1.826	379477	56515		1.15	2685
2	Nebivolol	4.443	287871	15973	5.91	1.44	2269

Accuracy

Table 2: Results of Accuracy for sample concentration-80%

	Name	RT	Area	Height	USP Resolution	USP Tailing	USP Plate Count	Injection
1	Ivabradine	1.822	2643.26	38605		1.19	2672	1
2	Nebivolol	4.434	222044	12431	8.25	1.59	2201	1
3	Ivabradine	1.824	2575.66	37548		1.17	2583	2
4	Nebivolol	4.436	222571	11819	7.58	1.62	2910	2
5	Ivabradine	1.818	2621.84	38105		1.21	2515	3
6	Nebivolol	4.426	224674	12405	8.61	1.60	2192	3




Fig. 1: Chromatogram for sample concentration-80%

Linearity

Table 3: Results of method linearity for Nebivolol

S.No	Peak Name	RT	Area (µV*sec)	Height (µV)	USP Plate Count	USP Tailing	USP Resolution
1	Nebivolol	4.444	41922	3260	3045.7	1.2	9.8
2	Nebivolol	4.430	141840	8019	2262.7	1.6	8.7
3	Nebivolol	4.432	275422	14434	2617.2	1.6	6.8
4	Nebivolol	4.435	307888	15858	2311.3	1.5	6.3
5	Nebivolol	4.435	384272	20111	2592.7	1.6	6.7
<b>Mean</b>			<b>230268.8</b>		2965.9	1.5	7.6
<b>Std. Dev.</b>			<b>136998.7</b>				
<b>% RSD</b>			<b>59.5</b>				




Fig. 2: Calibration curve of Ivabradine and Nebivolol

Table 4: Results of method linearity for Ivabradine

S.No	Peak Name	RT	Area (µV*sec)	Height (µV)	USP Plate Count	USP Tailing
1	Ivabradine	1.819	107480	14701	2440.0	1.2
2	Ivabradine	1.820	212590	31050	2615.4	1.2
3	Ivabradine	1.822	385507	54259	2582.4	1.2
4	Ivabradine	1.822	417339	56591	2406.0	1.1
5	Ivabradine	1.821	518455	75642	2617.1	1.2
<b>Mean</b>			<b>328274.1</b>		2532.2	1.2
<b>Std. Dev.</b>			<b>165461.6</b>			
<b>% RSD</b>			<b>50.4</b>			

Table 5: Results of LOQ

	Name	RT	Area	Height	USP Resolution	USP Tailing	USP Plate Count	Injection
1	Ivabradine	1.826	364591	52312		1.13	2596	3
2	Nebivolol	4.465	258472	14772	6.46	1.67	2299	3

Intermediate Precision

Table 6: Results of method precession for Ivabradine

S. No	Peak name	Retention time	Area(µV*sec)	Height(µV)	USP Plate Count	USP Tailing
1	Ivabradine	1.825	377285	55373	2569.8	1.2
2	Ivabradine	1.824	379768	56287	2691.8	1.2
3	Ivabradine	1.826	380712	56033	2632.8	1.2
4	Ivabradine	1.822	382966	55330	2592.1	1.2
5	Ivabradine	1.822	388290	57003	2668.9	1.2
<b>mean</b>			<b>381804.2</b>		2631.1	1.2
<b>Std.dev</b>			<b>4158.5</b>			
<b>%RSD</b>			<b>1.1</b>			

Ruggedness

Table 7: Results of method ruggedness for Ivabradine

S.No	Peak Name	RT	Area (µV*sec)	Height (µV)	USP Plate Count	USP Tailing
1	Ivabradine	1.827	380388	55093	2653.9	1.2
2	Ivabradine	1.819	384418	53697	2397.9	1.2
3	Ivabradine	1.820	388201	53851	2442.2	1.2
<b>Mean</b>			<b>384335.6</b>		2498.0	1.2
<b>Std. Dev.</b>			<b>3907.5</b>			
<b>% RSD</b>			<b>1.0</b>			

**Table 8: Results of method Ruggedness for Nebivolol**

S.No	Peak Name	RT	Area ( $\mu\text{V}\cdot\text{sec}$ )	Height ( $\mu\text{V}$ )	USP Plate Count	USP Tailing	USP Resolution
1	Nebivolol	4.418	286892	15040	2833.4	1.5	7.3
2	Nebivolol	4.433	288548	15833	2184.3	1.6	8.3
<b>Mean</b>			<b>287831.6</b>		2009.7	1.6	7.8
<b>Std. Dev.</b>			<b>850.3</b>				
<b>% RSD</b>			<b>0.3</b>				

**CONCLUSION**

From the above experimental results and parameters it was concluded that, this newly developed method for the simultaneous estimation of Ivabradine and Nebivolol in Tablet Dosage Form was found to be simple, precise, accurate and high resolution and shorter retention time makes this method more acceptable and cost effective and it can be effectively applied for routine analysis in research institutions, quality control department in meant in industries, approved testing laboratories, biopharmaceutical and bio-equivalence studies and in clinical pharmacokinetic studies in near future.

**REFERENCES**

1. Sharma BK. Instrumental method of chemical analysis, 23<sup>rd</sup> edition, goa publishers. 2004.
2. Lloyd R Snyder, Joseph J Kirkland and Joseph L Glajch. Practical HPLC method development. second edition, 1, 420-430,686-704.
3. Becket and steelmaker, practical pharmaceutical chemistry, part 24<sup>th</sup> edition CBS publications and distributors. 2005;157-168.
4. Seth PD. HPLC quantitative analysis of pharmaceutical formulations CBS publications and distributors. 1<sup>st</sup> edition, 2001;69-70.
5. Validating chromatographic methods, David M.Bliesner. 1-4.
6. Irving Claire B, Adams Clive E, Lawrie and Stephen. Haloperidol versus Placebo for SACHizophrenia. In Irving, Claire B. Cochrane Database of Systematic Reviews. 2006;(4): CD003082.
7. International conference on harmonization: ICH Q 2 (R1) Validation of Analytical Procedures: Text and Methodology. 1995.
8. Indian pharmacopeia. 2007;1:715.
9. British pharmacopeia. 2007;1:136.
10. Martindale the complete drug reference, thirty sixth editions.